

Mini Implant Mandibular Overdenture Attachments

Last Updated Friday, 25 August 2006

Immediate Load Mini Implants For a Mandibular Overdenture

The attachment caps are placed on the implants. They contain rubber O-rings which are the active part of the attachment

The metal caps are picked up in the denture using acrylic resin. Excess resin is cleared away and the prosthesis is complete. The O-rings can be easily changed at any time.